



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center – WO66-0609
Silver Spring, MD 20993-0002

SEP 14 2010

The Binding Site Group, Ltd
c/o Mr. Jay H. Geller
Authorized U.S. Representative
12100 Wilshire Boulevard, Suite 500
Los Angeles, CA 90025-7121

Re: k100179

Trade/Device Name: Human C3c Kit for use on SPAPLUS™
Regulation Number: 21 CFR §866.5240
Regulation Name: Complement components immunological test system
Regulatory Class: Class II
Product Code: CZW
Dated: August 2, 2010
Received: August 9, 2009

Dear Mr. Geller:

This letter corrects our substantially equivalent letter of 8/13/2009.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

Page 2 – Mr. Jay H. Geller

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Maria M. Chan, Ph.D.
Director
Division of Immunology and Hematology Devices
Office of *In Vitro* Diagnostic Device Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

K100179

Indications for Use

AUG 13 2010

510(k) Number (if known): K100179

Device Name: Human C3c Kit for use on SPAPLUS™

Indications for Use: This kit is intended for the quantitative *in vitro* determination of human C3c in serum using the Binding Site SPAPLUS™ turbidimetric analyser. This test should be used in conjunction with other laboratory and clinical findings.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Mandi M Chan Attachment D
Division Sign-Off

Page 1 of 1

Office of In Vitro Diagnostic
Device Evaluation and Safety

510(k) K100179

14